



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Northeast Metrology, Inc.**

**140 Industrial Drive**

**East Longmeadow, MA 01028**

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 15 February 2027

Certificate Number: AC-1519



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

**Northeast Metrology, Inc.**

140 Industrial Drive  
East Longmeadow, MA 01028  
Mark Kuehl 413-525-1502

### CALIBRATION

Valid to: **February 15, 2027**

Certificate Number: **AC-1519**

#### Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 4 in (5 to 20) in	(1.8 + 3L) μin (3 + 3.2L) μin	Comparison to Gage Blocks
Regular and Thread Micrometer Standards <sup>2</sup>	(1 to 20) in (21 to 72) in	(1 + 3.7L) μin (130 + 3.5L) μin	Comparison to Universal Measuring Machine (UMM), Gage Blocks
Flute, O.D., Depth, Interchangeable-Anvil Micrometers <sup>1,2</sup>	Up to 72 in	(100 + 4.2L) μin	Comparison to Gage Blocks
Calipers <sup>1,2</sup> (Dial, Digital, Vernier)	Up to 120 in	(570 + 1.4L) μin	Comparison to Gage Blocks, Ring Gage
Pitch/Gear Wire Sets (Up to 130 TPI)	Up to 0.25 in	14 μin	Comparison to UMM, Pin Gages
Thread Plugs <sup>2</sup> Pitch Diameter	Up to 12 in	(66 + 2.6L) μin	Comparison to UMM, Gage Blocks, Pitch Wires
Major Diameter	Up to 12 in	(27 + 3.2L) μin	
Thread Rings <sup>2,3</sup> Pitch Diameter	Up to 6 in	(90 + 2.8L) μin	Comparison to Master Thread Plugs
Minor Diameter	Up to 6 in	(630 + 1.3L) μin	

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Plain Plugs/Discs <sup>2</sup>	(0.005 to 12) in	$(9 + 3.9L) \mu\text{in}$	Comparison to UMM, Gage Blocks
Plain Ring Gages <sup>2</sup>	(0.04 to 10) in	$(8 + 4L) \mu\text{in}$	Comparison to Ring/Disc Comparator, Gage Blocks
Electronic, Dial, Test Indicators <sup>1,2</sup>	(0.000 05 to 4) in	$(39 + 3.5L) \mu\text{in}$	Comparison to Indicator Calibrator
Height Gages <sup>1,2</sup>	Up to 24 in	$(123 + 3.5L) \mu\text{in}$	Comparison to Gage Blocks, Surface Plate
Pin Gages <sup>2</sup>	(0.011 to 1) in	$(15 + 5.6L) \mu\text{in}$	Comparison to UMM, Reference Pin Gages, Laser Micrometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

### Notes:

- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
- $L$  = length in inches.
- Only the Minor Diameter measurement is measured and Accredited. The Thread Set Plug is utilized is for tactical fit. If the ring gage does not fit correctly, it is then adjusted to the Thread Set Plug and the Minor Diameter will be rechecked. If proper drag is felt, if it is loose, or if it is tight, it will be checked on the certificate.
- Unless otherwise specified in the far-right column, the calibration procedure/method was internally written.
- This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1519.



Jason Stine, Vice President